



Schweizerische Eidgenossenschaft

Confédération suisse

Confederazione Svizzera

Confederaziun svizra

Swiss Confederation

Federal Department of Economic Affairs,
Education and Research EAER

State Secretariat for Economic Affairs SECO

Swiss Accreditation Service SAS

SCS Directory

Accreditation number: SCS 0108

International standard: ISO/IEC 17025:2017

Swiss standard: SN EN ISO/IEC 17025:2018

Schmid & Partner Engineering AG
Zeughausstrasse 43
8004 Zürich
(Switzerland)

Head: Dr Sven Kühn
Responsible for MS: Prof. Dr Niels Kuster
Telephone: +41 44 245 97 00
E-Mail: info@speag.swiss
Internet: <http://www.speag.swiss>
Initial accreditation: 17.09.2004
Current accreditation: 17.09.2019 to 16.09.2024
Scope of accreditation see: www.sas.admin.ch
(Accredited bodies)

Scope of accreditation as of 12.05.2023

Calibration Laboratory for Electric Field, Magnetic Field and SAR Sensors and Dielectric Measurement Instrumentation

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Electric field				
Calibration of E-field probes	0.8 V/m ... 800 V/m	4 MHz ... 6 GHz	5,1 %	e.g. ER3DV6x, EF3DVx, EU2DVx, EE3DVx, EL3DVx
Calibration of E-field probes	10 V/m ... 2000 V/m	750 MHz ... 6 GHz 6 GHz ... 110 GHz	5,1 % 0,98 dB	e.g. EUmmWVx
Calibration of E-field probes	2 V/m ... 420 V/m	3 kHz ... 10 MHz	1,06 dB	e.g., MAGPy-8H3D+E3D
Magnetic field				
Calibration of H-field probes	2 mA/m ... 2 A/m	4 MHz ... 3 GHz	5,1 %	e.g. H2DVx, H3DVx, HL3DVx
Calibration of H-field probes	10 A/m ... 2200 A/m	3 kHz ... 10 MHz	0,70 dB	e.g., MAGPy-8H3D+E3D



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration of sensitivity for magnetic field probes in the audio range	0,001 ... 0,1 V/(A/m)	1 kHz 0,1 ... 1 A/m	2,2 %	e.g. AM1DVx
Calibration of magnetic field simulator	-30 ... +40 dB A/m	1 kHz	4,1 %	e.g. TMFS (Telephone Magnetic Field Simulator)
Specific absorption rate (SAR)	E* field (typical ²)			e.g. ET3Dvx, ES3Dvx, EX3Dvx, ET1Dvx, EU2Dvx
Calibration of dosimetric E-field probes	0,5 V/m ... 500 V/m	4 MHz ... 450 MHz	6,7 % (13,3 % for SAR)	Temperature transfer calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 450 MHz
	E* field (typical ²)			e.g. ET3Dvx, ES3Dvx, EX3Dvx, ET1Dvx, EU2Dvx
Calibration of dosimetric E-field probes	0,45 V/m ... 450 V/m	750 MHz ... 3 GHz	5,5 % (11 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 1800 MHz
	0,4 V/m ... 450 V/m	3 GHz ... 6 GHz	6,5 % (13,1 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 5200 MHz
	0,4 V/m ... 400 V/m	6 GHz ... 10 GHz	9,3 % (18,6 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 7 GHz



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration of temperature SAR probes	E* field (typical) ²⁾ 0 °C ... + 60 °C	Tissue simulating Liquids	0,15 K (5 % temperature gradient for SAR)	e.g. EX3DVx, ET1DVx As example, the temperature gradient of T1Vx and T1V3LAB probes can be determined to 5 %, which is also contribution to SAR accuracy. (Noise is dominating the lower SAR threshold to typically 0,2 W/kg)
Calibration of test system validation sources	SAR* 1 g and 10 g per 1 W input power	4 MHz ... 290 MHz 300 MHz ... 700 MHz 700 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 10 GHz	18,4 % for SAR 1 g 18,0 % for SAR 10 g 18,1 % for SAR 1 g 17,6 % for SAR 10 g 17,0 % for SAR 1 g 16,5 % for SAR 10 g 19,9 % for SAR 1 g 19,5 % for SAR 10 g 24,7 % for SAR 1 g 24,4 % for SAR 10 g	e.g. CLA-6, CLA-13, CLA-30, CLA-64, CLA-128, CLA-150, CLA-220 e.g. D835V2 ... D3000V2 according to IEEE 1528, IEC 62209-1/2, for 1 g and 10 g SAR e.g. D3500V2 ... D5GHzV2 according to IEC 62209-2, for 1 g and 10 g SAR e.g. D6.5GHzV2 ... D9GHzV2 according to IEC 62209-U, for 1 g and 10 g SAR *) SAR given (as example) for head tissue simulating liquid *) SAR given (as example) for head tissue simulating liquid
Near-Field Calibration of dipoles in air	E* field per 0,1 W input power 30 V/m ... 300 V/m	700 MHz ... 6 GHz	9,5 % for E field	e.g. CD835V3 ... CD5500V3 according to ANSI C 63.19, for E field and H field
	H* field per 0,1 W input power 0,07 A/m ... 0,7 A/m	385 MHz – 3 GHz	8,3 % for H field	e.g. AdxxxV5 according to ANSI 14117 for H-field
Calibration of Golden Validation Device	SAR* 1 g and 10 g per GVD input power	900 MHz	15,0 % for SAR 1 g 15,1 % for SAR 10 g	e.g. GVD – SR 004001 AA



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
DC Voltage				
Calibration of readout units for field and SAR probes	2 mV 200 mV 1 mV ... 500 mV		1,5 % 1 % 1,16 %	e.g. DAE3Vx, DAE4Vx, DAEeasyVx
Calibration of SAR for planar array systems	SAR* peak at 4 mm depth per 1 W input power SAR* peak at 3 mm depth per 1 W input power SAR* peak at 3 mm depth per 1 W input power SAR* peak at 3 mm depth per 1 W input power	650 MHz ... 3 GHz 650 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 10 GHz	25,8 % for iSAR Flat 29,3 % for iSAR Head 22,4 % for cSAR3D Flat 25,9 % for cSAR3D Left/Right Head 25,1 % for cSAR3D Flat 28,3 % for cSAR3D Left/Right Head 29,8 % for cSAR3D Flat	e.g. iSAR Flat, iSAR Head e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head e.g. cSAR3D Flat, cSAR3D Quad
Calibration of antenna sources for transfer calibration of planar array systems	SAR* peak per 1 W input power SAR* peak per 1 W input power	650 MHz ... 3 GHz 3 GHz ... 6 GHz	17,7 % for SAR peak 19,9 % for SAR peak	e.g., SA AAE 083B, SA AAE 190 A, etc.
Calibration of thermometers	0 °C ... + 60 °C		0,1K	
Calibration of active electro-optical E&H-field probes optimized for close near-field evaluations in air	15 mV/m ... 75 V/m 42 µA/m ... 0,5 A/m	50 MHz ... 6000 MHz 50 MHz ... 6000 MHz	3,3 dB 3,3 dB	for antenna factor (E-field) for antenna factor (H-field)
Calibration of active electro-optical RF over fiber systems	Gain: -55 ... 15dB	10 MHz ... 10 GHz	2,0 dB	For transfer function (gain)
Calibration of stabilized RF power sources	- 5 dBm ... + 17 dBm	600 MHz ... 6 GHz	0,43 dB	e.g. Powersource1
Calibration Procedure for sources in air above 6 GHz	10 V/m ... 2000 V/m	10 GHz... 110 GHz	1,27 dB 1,47 dB 1,54 dB	e.g. verification source e.g. validation source e.g. general source



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration Procedure for sources in air above 6 GHz, for H and averaged S	25 mA/m ... 2,5 A/m 2.5 W/m ² ...2,5 kW/m ²	10 GHz... 110 GHz	1,28 dB 1,63 dB 1,70 dB	e.g. verification source e.g. validation source e.g. general source
Measurement capability of the dielectric probe	Permittivity, Conductivity or Loss Tangent from 4 MHz ... 67 GHz	homogeneous isotropic material		Open-ended coaxial probes, e.g. DAK-12, DAK-3.5, DAK-1.2 E
Permittivity	1...15		4 MHz ... 20 MHz 20 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz 40 GHz ... 67 GHz	Loss tangent < 0,1 Conductivity: > 10 S/m
	10...40		4 MHz ... 10 MHz 10 MHz ... 50 MHz 50 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	Conductivity < 0,1 S/m Conductivity < 0,1 S/m Conductivity: 0,1 – 10 S/m Conductivity: 0,1 – 10 S/m Conductivity: 1 – 10 S/m Conductivity: > 10 S/m Conductivity: > 10 S/m Conductivity: > 10 S/m
	35...100		4 MHz ... 10 MHz 10 MHz ... 50 MHz 50 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz	Conductivity: 0,1 – 1 S/m Conductivity: 0,1 – 1 S/m Conductivity: 1 – 10 S/m Conductivity: 1 – 10 S/m Conductivity: > 10 S/m
Conductivity	1 ... 10 S/m		4 MHz ... 10 MHz 10 MHz ... 50 MHz 50 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz	Permittivity: 35 – 100 Permittivity: 10 – 40
	>10 S/m		4 MHz ... 10 MHz 10 MHz ... 50 MHz 50 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	Permittivity: 10 – 40 Permittivity: 10 – 40 Permittivity: 1 – 15



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Loss Tangent	0 ... 0,1			
		4 MHz ... 20 MHz	0,46	
		20 MHz ... 200 MHz	0,28	
		200 MHz ... 3 GHz	0,03	
		3 GHz ... 6 GHz	0,03	
		6 GHz ... 20 GHz	0,03	
		20 GHz ... 40 GHz	0,03	
		40 GHz ... 67 GHz	0,03	
Measurement capability of the dielectric probe for liquids and gels	Permittivity and Conductivity from 4 MHz ... 67GHz	homogeneous isotropic material		Open-ended coaxial probes, e.g. DAK-12, DAK-3.5, DAK-1.2 E
Permittivity	1...100			
		4 MHz ... 10 MHz	6,4 %	
		10 MHz ... 20 MHz	3,8 %	static conductivity < 0,1 S/m
		20 MHz ... 30 MHz	2,3 %	
		30 MHz ... 50 MHz	1,9 %	
		50 MHz ... 5 GHz	2,2 %	
		5 GHz ... 20 GHz	3,7 %	
	1...100			static conductivity 0,1 – 10 S/m
		4 MHz ... 10 MHz	6,8 %	
		10 MHz ... 20 MHz	3,2 %	
		20 MHz ... 30 MHz	3,3 %	
		30 MHz ... 50 MHz	3,3 %	
		50 MHz ... 5 GHz	3,1 %	
		5 GHz ... 20 GHz	3,9 %	
		20 GHz ... 40 GHz	5,0 %	
		40 GHz ... 67 GHz	6,5 %	
Conductivity	0,01 ... 0,1 S/m			Permittivity: 1 – 100
		300 MHz ... 500 MHz	7,5 %	
		500 MHz ... 5 GHz	4,4 %	
		5 GHz ... 20 GHz	5,4 %	
Conductivity	0,1 – 10 S/m			Permittivity: 1 – 100
		4 MHz ... 10 MHz	3,9 %	
		10 MHz ... 20 MHz	3,9 %	
		20 MHz ... 30 MHz	3,8 %	
		30 MHz ... 50 MHz	3,4 %	
		50 MHz ... 5 GHz	4,8 %	
		5 GHz ... 20 GHz	4,8 %	
		20 GHz ... 40 GHz	4,8 %	
		40 GHz ... 67 GHz	4,9 %	



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
Capability of the dielectric probe to measure thin layers of solids and liquids in small volume	Permittivity, Conductivity or Loss Tangent from 4 MHz ... 67 GHz	Homogeneous isotropic material	Permittivity	Loss tangent	Open-ended coaxial probes, e.g. DAK-TL-12, DAKTL-3.5, DAKTL-1.2 E
Permittivity	1 < Permittivity < 10	4 MHz ... 20 MHz	---	---	0.1 < thickness < 0.2 mm
	Loss tangent < 0,05	20 MHz ... 30 MHz	---	---	
		30 MHz ... 50 MHz	---	---	
		50 MHz ... 100 MHz	---	---	
		100 MHz ... 600 MHz	32,6 %	0,06	
		600 MHz ... 3 GHz	29,5 %	0,03	
		3 GHz ... 6 GHz	12,6 %	0,03	
		6 GHz ... 20 GHz	10,0 %	0,03	
		20 GHz ... 40 GHz	9,1 %	0,03	
		40 GHz ... 67 GHz	4,5 %	0,03	
Permittivity	1 < Permittivity < 10	4 MHz ... 20 MHz	34,6 %	0,45	0.2 < thickness < 1 mm
	Loss tangent < 0,05	20 MHz ... 30 MHz	27,0 %	0,27	
		30 MHz ... 50 MHz	25,6 %	0,17	
		50 MHz ... 100 MHz	20,7 %	0,10	
		100 MHz ... 600 MHz	9,1 %	0,06	
		600 MHz ... 3 GHz	6,5 %	0,03	
		3 GHz ... 6 GHz	3,7 %	0,03	
		6 GHz ... 20 GHz	3,3 %	0,03	
		20 GHz ... 40 GHz	3,9 %	0,03	
		40 GHz ... 67 GHz	3,5 %	0,03	
Permittivity	1 < Permittivity < 10	4 MHz ... 20 MHz	24,3 %	0,45	1 < thickness < 10 mm
	Loss tangent < 0,05	20 MHz ... 30 MHz	11,2 %	0,27	
		30 MHz ... 50 MHz	7,1 %	0,17	
		50 MHz ... 100 MHz	4,7 %	0,10	
		100 MHz ... 600 MHz	2,7 %	0,06	
		600 MHz ... 3 GHz	2,1 %	0,03	
		3 GHz ... 6 GHz	2,0 %	0,03	
		6 GHz ... 20 GHz	2,2 %	0,03	
		20 GHz ... 40 GHz	3,9 %	0,03	
		40 GHz ... 67 GHz	3,2 %	0,03	
Permittivity	1 < Permittivity < 10	4 MHz ... 20 MHz	---	---	0.1 < thickness < 0.2 mm
	Loss tangent > 0,05	20 MHz ... 30 MHz	---	---	
		30 MHz ... 50 MHz	---	---	
		50 MHz ... 100 MHz	---	---	
		100 MHz ... 600 MHz	18,6 %	0,06	
		600 MHz ... 3 GHz	24,6 %	0,03	
		3 GHz ... 6 GHz	14,7 %	0,03	
		6 GHz ... 20 GHz	8,3 %	0,03	
		20 GHz ... 40 GHz	10,4 %	0,03	
		40 GHz ... 67 GHz	4,7 %	0,05	



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
	1< Permittivity<10 Loss tangent.> 0,05	4 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 100 MHz 100 MHz ... 600 MHz 600 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	--- --- --- --- 29,8 % 19,5 % 8,0 % 4,1 % 4,3 % 3,8 %	--- --- --- --- 0,06 0,03 0,03 0,03 0,03	0.2<thickness<1 mm
	1< Permittivity<10 Loss tangent.> 0,05	4 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 100 MHz 100 MHz ... 600 MHz 600 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	24,3 % 11,2 % 7,1% 4,7 % 2,6 % 2,0 % 1,9 % 2,0 % 3,2 % 3,1 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03	1 <thickness<10 mm
	Permittivity>10 Loss tangent< 0,05	4 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 100 MHz 100 MHz ... 600 MHz 600 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	28,4 % 18,5 % 12,6 % 8,6 % 5,7 % 5,7 % 5,7 % 4,1 % 4,6 % 6,2 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	0.1 <thickness <0.2 mm
	Permittivity>10 Loss tangent< 0,05	4 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 100 MHz 100 MHz ... 600 MHz 600 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	24,7 % 12,1 % 8,5 % 6,6 % 3,7 % 4,0 % 3,0 % 3,5 % 3,8 % 3,9 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	0.2<thickness<1 mm
	Permittivity>10 Loss tangent< 0,05	4 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 100 MHz 100 MHz ... 600 MHz 600 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz 20 GHz ... 40 GHz 40 GHz ... 67 GHz	24,3 % 11,2 % 7,1% 4,7 % 2,6 % 1,9 % 1,9 % 2,0 % 4,5 % 3,6 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	1 <thickness<10 mm



SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
Permittivity Loss tangent> 0,05	4 MHz ... 20 MHz	20,9 %	0,35		0.1 < thickness < 0.2 mm
	20 MHz ... 30 MHz	20,4 %	0,35		
	30 MHz ... 50 MHz	15,3 %	0,35		
	50 MHz ... 100 MHz	11,2 %	0,25		
	100 MHz ... 600 MHz	7,9 %	0,11		
	600 MHz ... 3 GHz	7,2 %	0,02		
	3 GHz ... 6 GHz	5,3 %	0,03		
	6 GHz ... 20 GHz	4,2 %	0,05		
	20 GHz ... 40 GHz	6,4 %	0,06		
	40 GHz ... 67 GHz	8,9 %	0,14		
Permittivity Loss tangent> 0,05	4 MHz ... 20 MHz	7,6 %	0,35		0.2 < thickness < 1 mm
	20 MHz ... 30 MHz	6,1 %	0,35		
	30 MHz ... 50 MHz	6,1 %	0,35		
	50 MHz ... 100 MHz	6,0 %	0,25		
	100 MHz ... 600 MHz	4,8 %	0,11		
	600 MHz ... 3 GHz	5,4 %	0,02		
	3 GHz ... 6 GHz	3,5 %	0,03		
	6 GHz ... 20 GHz	3,8 %	0,05		
	20 GHz ... 40 GHz	5,2 %	0,06		
	40 GHz ... 67 GHz	6,0 %	0,08		
Permittivity Loss tangent> 0,05	4 MHz ... 20 MHz	5,2 %	0,35		1 < thickness < 10 mm
	20 MHz ... 30 MHz	2,6 %	0,35		
	30 MHz ... 50 MHz	2,6 %	0,35		
	50 MHz ... 100 MHz	2,5 %	0,25		
	100 MHz ... 600 MHz	2,5 %	0,11		
	600 MHz ... 3 GHz	2,8 %	0,02		
	3 GHz ... 6 GHz	2,8 %	0,03		
	6 GHz ... 20 GHz	3,6 %	0,05		
	20 GHz ... 40 GHz	3,3 %	0,06		
	40 GHz ... 67 GHz	4,5 %	0,08		
Permittivity Conductivity	Permittivity		Conduc-	Permittivity	1 < thickness < 10 mm
	Conductivity > 0,5 S/m		tivity	Conduc-	
	4 MHz ... 20 MHz	6,4 %	3,9 %	3,9 %	
	20 MHz ... 30 MHz	4,3 %	3,4 %	3,4 %	
	30 MHz ... 50 MHz	4,0 %	3,4 %	3,4 %	
	50 MHz ... 100 MHz	2,5 %	3,4 %	3,4 %	
	100 MHz ... 600 MHz	2,9 %	3,4 %	3,4 %	
	600 MHz ... 3 GHz	2,8 %	5,8 %	5,8 %	
	3 GHz ... 6 GHz	2,8 %	4,0 %	4,0 %	
	6 GHz ... 20 GHz	3,6 %	4,0 %	4,0 %	
	20 GHz ... 40 GHz	3,9 %	4,1 %	4,1 %	
	40 GHz ... 67 GHz	5,1 %	5,0 %	5,0 %	

(2) Slightly depending on the frequency and probe type.

* / * / * / * / *